



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

0000 7813

DATE: 4/20/2006 INITIALS: DEC OWNER: WSP
DISTRICT: 2 ADDRESS: _____
LOCATION: 7th Ship COUNTY: _____
MAKE: General TYPE: port wheel weigher ACCURACY CLASS: III
MODEL: 101 ID. NO: 27602 CAPACITY: 20,000 lbs SECTIONS: 1
MINIMUM GRADUATIONS: 50 REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15000	0				
Center	20000	20050	+50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cramer Date: 4/26/2006

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before
me on 4-26-06 by

Michael J. Evans
(Signature of notary public)
Michael J. EVANS

Printed Name of Notary 12-27-07
My appointment expires



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS